152995 SCALER SYSTEM 240 VAC (ASEPSIS)


(1) 153192 SCALER ASS'Y, ASEPSIS
(1) 120661 INSERT, HANDPIECE
(1) 117481 FITTING
(1) 117477 TEE
GREEN TUBING
CLEAR TUBING
KINK VALVE
FLOW CONTROL KNOB
(3) 103152 RECEPTACLE
(3) 101949 SOC TERMINAL
(1) 153191 PRESSURE SWITCH ASS'Y
PRESSURE SWITCH TO BE MOUNTED IN INST. HEAD.
POTENTIOMETER HOLE FOR SCALER .275 DIA ON OPPOSITE SIDE OF PILOT VALVE
(1) 153189 POTENTIOMETER
(1) 120875-1 KNOB
(1) 120178 SET SCREW
(1) 117244 CONNECTOR
(1) 103167 RECEPTACLE
(2) 101949 SOC TERMINAL
(1) 115426 PLUG, 6 CIR
(5) 101928 PIN, TERMINAL
UMBILICAL FOR DRS CART & LOW COST UNITS WITHOUT AN ASS'Y CONSOLE. ADD PIN TERMINALS & PLUG TO END OF COLORED WIRES AS SHOWN.
RED 1 4 BROWN
GREEN 2 5 BLACK
WHITE 3 6
(1) 153190 CABLE, SCALER UMBILICAL NOTE: AFTER CABLE & PLUG HAVE BEEN ADDED TO UMBILICAL, CABLE TO BE 100% TESTED IN TESTER RT10919.
(1) 153188 SCALER/SELONOID ASS'Y 24V
NOTE: TUBING CONNECTING FOR A SELF-CONTAINED WATER SYSTEM (REFERENCE ONLY)
TIN
(.34 FT) R5887 WIRE, BLUE
.50
4.00
(1) 121771 TERMINAL
UMBILICAL FOR ASS'T CONSOLE & CONCEPT L/R
Ⓐ UMBILICAL DUO CART AND DUO WALL MOUNT ADD 9' 177523 BLUE 1/8" OD TUBING
BLUE TUBING
(4) P5788 SCREW
(4) 122357 NUT
RUN SCREW FROM BOTTOM OF APC, NUT ON TOP
(1) 153045 TRANSFORMER ASS'Y, 240V
BLUE
BROWN
(1) 122541 FUSE HOLDER
(1) 120497 FUSE, 1/2 AMP
NOTE: TEST APC WITH RT10923 TESTER
NOTE: TUBING CONNECTING FOR CITY WATER IN JUNCTION BOX (REFERENCE ONLY)
(1) 153188 SELONOID (REF)
(1 FT/12 IN) 177523 BLUE TUBING
BLUE TUBING & TEE TO BE ADDED TO SCALER ONLY FOR CITY WATER.
(1) 115426 PLUG
Ⓐ(5) 101928 PIN TERMINAL
AIR REGULATOR IN
WATER REGULATOR IN
(1) 117477 TEE

NOTES: *USE P/N 119479 UNI-CLAMP WHERE APPLICABLE
*DEPENDING ON INSTALLATION, ALL THE PARTS WILL NOT BE USED. RETURN ALL UNUSED PARTS TO STOCK.
*ADD (1) TPW1716 INSTRUCTIONS TO KIT.

ENTIRE SYSTEM TO BE 100% TESTED